

JUL - 8 2011

510(K) SUMMARY

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of SMDA 1990 and 21 CFR §807.92(c).

The assigned 510(k) number is: K111435.

1. Submitter:

Shenzhen Mindray Bio-medical Electronics Co., LTD
Mindray Building, Keji 12th Road South, Hi-tech Industrial Park, Nanshan, Shenzhen,
518057, P. R. China

Tel: +86 755 8188 5635

Fax: +86 755 2658 2680

Contact Person:

Bai Yanhong

Shenzhen Mindray Bio-medical Electronics Co., LTD

Mindray Building, Keji 12th Road South, Hi-tech Industrial Park,
Nanshan, Shenzhen, 518057, P. R. China

Date Prepared: May 9, 2011

2. Device Name:

DP-50 Digital Ultrasonic Diagnostic Imaging System

Classification

Regulatory Class: II

21 CFR 892.1560 Ultrasonic Pulsed Echo Imaging System (90-IYO)

21 CFR 892.1570 Diagnostic Ultrasound Transducer (90-ITX)

3. Device Description:

The DP-50 Digital Ultrasonic Diagnostic Imaging System is a general purpose, mobile, software controlled, ultrasonic diagnostic system. Its function is to acquire and display ultrasound data in B-Mode, M-Mode, or their combined mode B+M Mode. This system is a Track 3 device that employs an array of probes that include linear array and convex array with a frequency range of approximately 3.5 MHz to 10.0 MHz.

4. Intended Use:

The DP-50 Digital Ultrasonic Diagnostic Imaging System is applicable for adults, pregnant women, pediatric patients and neonates. It is intended for use in fetal, abdominal, pediatric, small organ(breast, thyroid, testes), neonatal and adult cephalic, trans-rectal, trans-vaginal, musculo-skeletal(conventional, superficial), cardiac(pediatric) and peripheral vascular exams

5. Comparison with Predicate Device:

DP-50 Digital Ultrasonic Diagnostic Imaging System is comparable with and substantially equivalent to the Mindray DP-6900 Digital Ultrasonic Diagnostic Imaging System (K090912), M5 Diagnostic Ultrasound System(K102991) and M7 Diagnostic Ultrasound System(K103677). They have the same technological characteristics, are comparable in key safety and effectiveness features, and have the same intended uses and basic operating modes as the predicate device.

6. Non-clinical Tests:

DP-50 Digital Ultrasonic Diagnostic Imaging System has been evaluated for acoustic output, biocompatibility, cleaning and disinfection effectiveness as well as thermal, electrical and mechanical safety, and has been found to conform with applicable medical safety standards. This device has been designed to meet the following standards: UD 2, UD 3, IEC 60601-1, IEC 60601-1-1, IEC 60601-1-2, IEC 60601-2-37, IEC 60601-1-4, ISO 10993-1 and IEC 62304.

Conclusion:

Intended uses and other key features are consistent with traditional clinical practices, FDA guidelines and established methods of patient examination. The design, development and quality process of the manufacturer confirms with 21 CFR 820, ISO 9001 and ISO 13485 quality systems. The device conforms to applicable medical device safety standards. Therefore, the DP-50 Digital Ultrasonic Diagnostic Imaging System is substantially equivalent with respect to safety and effectiveness to devices currently cleared for market.



Food and Drug Administration
10903 New Hampshire Avenue
Silver Spring, MD 20993

Shenzhen Mindray Bio-Medical Electronics Co., Ltd.
% Ms. Susan D. Goldstein-Falk
Official Correspondent
mdi Consultants, Inc.
55 Northern Blvd., Suite 200
GREAT NECK NY 11021

JUL - 8 2011

Re: K111435
Trade/Device Name: DP-50 Digital Ultrasonic Diagnostic Imaging System
Regulation Number: 21 CFR 892.1560
Regulation Name: Ultrasonic pulsed echo imaging system
Regulatory Class: II
Product Code: IYO and ITX
Dated: June 15, 2011
Received: June 16, 2011

Dear Ms. Goldstein-Falk:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

This determination of substantial equivalence applies to the following transducers intended for use with the DP-50 Digital Ultrasonic Diagnostic Imaging System, as described in your premarket notification:

Transducer Model Number

35C50EA
65C15EA
65EC10EA
75L38EA

75L53EA
10L24EA
65EB10EA

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus permits your device to proceed to market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please go to <http://www.fda.gov/AboutFDA/CentersOffices/CDRH/CDRHOices/ucm115809.htm> for the Center for Devices and Radiological Health's (CDRH's) Office of Compliance. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

If you have any questions regarding the content of this letter, please contact Shahram Vaezy at (301) 796-6242.

Sincerely Yours,



Mary S. Pastel, Sc.D.
Director
Division of Radiological Devices
Office of In Vitro Diagnostic Device
Evaluation and Safety
Center for Devices and Radiological Health

Enclosure(s)

Indications for Use

510(k) Number (if known):

Device Name: DP-50 Digital Ultrasonic Diagnostic Imaging System

Indications For Use:

The DP-50 Digital Ultrasonic Diagnostic Imaging System is applicable for adults, pregnant women, pediatric patients and neonates. It is intended for use in fetal, abdominal, pediatric, small organ(breast, thyroid, testes), neonatal and adult cephalic, trans-rectal, trans-vaginal, musculo-skeletal(conventional, superficial), cardiac(pediatric) and peripheral vascular exams.


Prescription Use X
(Part 21 CFR 801 Subpart D)

AND/OR Over-The-Counter Use
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of In Vitro Diagnostic Devices (OIVD)

Page 1 of 1


(Division Sign-Off)
Division of Radiological Devices
Office of In Vitro Diagnostic Device Evaluation and Safety
510K K111435

0028

Diagnostic Ultrasound Indications for Use Form

System: DP-50 Digital Ultrasonic Diagnostic Imaging System
 Transducer: N/A
 Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation							
	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic								
Fetal	N	N					N	Note 1,2
Abdominal	N	N					N	Note 1,2
Intraoperative (specify)*								
Intraoperative (Neuro)								
Laparoscopic								
Pediatric	N	N					N	Note 1, 2
Small organ(specify)**	N	N					N	Note 1,2
Neonatal Cephalic	N	N					N	Note 1, 2
Adult Cephalic	N	N					N	Note 1, 2
Trans-rectal	N	N					N	Note 1, 2
Trans-vaginal	N	N					N	Note 1, 2
Trans-urethral								
Trans-esoph.(non-Card.)								
Musculo-skeletal Conventional	N	N					N	Note 1,2
Musculo-skeletal Superficial	N	N					N	Note 1,2
Intravascular								
Cardiac Adult								
Cardiac Pediatric	N	N					N	Note 1,2
Intravascular (Cardiac)								
Trans-esoph.(Cardiac)								
Intra-Cardiac								
Peripheral Vascular	N	N					N	Note 1, 2
Other (specify)***								

N=new indication; P=previously cleared by FDA; E=added under Appendix E

Additional comments: Combined modes: B+M.

*Intraoperative includes abdominal, thoracic, and vascular

**Small organ-breast, thyroid, testes

***Other use includes Urology.


Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: Biopsy Guidance

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)


 (Division Sign-Off)
 Division of Radiological Devices
 Office of In Vitro Diagnostic Device Evaluation and Safety
 510K K111435

0029

Diagnostic Ultrasound Indications for Use Form

System: DP-50 Digital Ultrasonic Diagnostic Imaging System
 Transducer: 35C50EA
 Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation							
	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic								
Fetal	P	P					P	Note 1, 2
Abdominal	P	P					P	Note 1, 2
Intraoperative (specify)*								
Intraoperative (Neuro)								
Laparoscopic								
Pediatric	P	P					P	Note 1, 2
Small organ(specify)**								
Neonatal Cephalic								
Adult Cephalic								
Trans-rectal								
Trans-vaginal								
Trans-urethral								
Trans-esoph.(non-Card.)								
Musculo-skeletal Conventional	P	P					P	Note 1, 2
Musculo-skeletal Superficial								
Intravascular								
Cardiac Adult								
Cardiac Pediatric								
Intravascular (Cardiac)								
Trans-esoph.(Cardiac)								
Intra-Cardiac								
Peripheral Vascular	P	P					P	Note 1, 2
Other (specify)***								

N=new indication; P=previously cleared by FDA; E=added under Appendix E

Additional comments: Combined modes: B+M.

*Intraoperative includes abdominal, thoracic, and vascular

**Small organ-breast, thyroid, testes

***Other use includes Urology.


Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: Biopsy Guidance

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)


 (Division Sign-Off)
 Division of Radiological Devices
 Office of In Vitro Diagnostic Device Evaluation and Safety
 510K K111435

0030

Diagnostic Ultrasound Indications for Use Form

System: DP-50 Digital Ultrasonic Diagnostic Imaging System
 Transducer: 65C15EA
 Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation							
	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic								
Fetal								
Abdominal	P	P					P	Note 1, 2
Intraoperative (specify)*								
Intraoperative (Neuro)								
Laparoscopic								
Pediatric	P	P					P	Note 1, 2
Small organ(specify)**								
Neonatal Cephalic	P	P					P	Note 1, 2
Adult Cephalic	P	P					P	Note 1, 2
Trans-rectal								
Trans-vaginal								
Trans-urethral								
Trans-esoph.(non-Card.)								
Musculo-skeletal								
Conventional								
Musculo-skeletal								
Superficial								
Intravascular								
Cardiac Adult								
Cardiac Pediatric	P	P					P	Note 1, 2
Intravascular (Cardiac)								
Trans-esoph.(Cardiac)								
Intra-Cardiac								
Peripheral Vascular								
Other (specify)***								

N=new indication; P=previously cleared by FDA; E=added under Appendix E

Additional comments: Combined modes: B+M.

*Intraoperative includes abdominal, thoracic, and vascular

**Small organ-breast, thyroid, testes

***Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: Biopsy Guidance

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)

Mary S. Patel

(Division Sign-Off)

Division of Radiological Devices

Office of In Vitro Diagnostic Device Evaluation and Safety

510K

K111435

0031

Diagnostic Ultrasound Indications for Use Form

System: DP-50 Digital Ultrasonic Diagnostic Imaging System
 Transducer: 65EC10EA
 Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation							Other (specify)
	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	
Ophthalmic								
Fetal	P	P					P	Note 1, 2
Abdominal								
Intraoperative (specify)*								
Intraoperative (Neuro)								
Laparoscopic								
Pediatric								
Small organ(specify)**								
Neonatal Cephalic								
Adult Cephalic								
Trans-rectal	P	P					P	Note 1, 2
Trans-vaginal	P	P					P	Note 1, 2
Trans-urethral								
Trans-esoph.(non-Card.)								
Musculo-skeletal Conventional								
Musculo-skeletal Superficial								
Intravascular								
Cardiac Adult								
Cardiac Pediatric								
Intravascular (Cardiac)								
Trans-esoph.(Cardiac)								
Intra-Cardiac								
Peripheral Vascular								
Other (specify)***								

N=new indication; P=previously cleared by FDA; E=added under Appendix E

Additional comments: Combined modes: B+M.

*Intraoperative includes abdominal, thoracic, and vascular

**Small organ-breast, thyroid, testes

***Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: Biopsy Guidance

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)


 (Division Sign-Off)

Division of Radiological Devices
 Office of In Vitro Diagnostic Device Evaluation and Safety

510K

K111435

0032

Diagnostic Ultrasound Indications for Use Form

System: DP-50 Digital Ultrasonic Diagnostic Imaging System
 Transducer: 75L38EA
 Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation							
	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic								
Fetal								
Abdominal	P	P					P	Note 1,2
Intraoperative (specify)*								
Intraoperative (Neuro)								
Laparoscopic								
Pediatric	P	P					P	Note 1,2
Small organ(specify)**	P	P					P	Note 1,2
Neonatal Cephalic	P	P					P	Note 1,2
Adult Cephalic								
Trans-rectal								
Trans-vaginal								
Trans-urethral								
Trans-esoph.(non-Card.)								
Musculo-skeletal Conventional	P	P					P	Note 1,2
Musculo-skeletal Superficial	P	P					P	Note 1,2
Intravascular								
Cardiac Adult								
Cardiac Pediatric								
Intravascular (Cardiac)								
Trans-esoph.(Cardiac)								
Intra-Cardiac								
Peripheral Vascular	P	P					P	Note 1,2
Other (specify)***								

N=new indication; P=previously cleared by FDA; E=added under Appendix E

Additional comments: Combined modes: B+M.

*Intraoperative includes abdominal, thoracic, and vascular

**Small organ-breast, thyroid, testes

***Other use includes Urology.

Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: Biopsy Guidance

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)


 (Division Sign-Off)

Division of Radiological Devices
 Office of In Vitro Diagnostic Device Evaluation and Safety

610K K41435

0033

Diagnostic Ultrasound Indications for Use Form

System: DP-50 Digital Ultrasonic Diagnostic Imaging System
 Transducer: 75L53EA
 Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation							Other (specify)
	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	
Ophthalmic								
Fetal								
Abdominal	P	P					P	Note 1,2
Intraoperative (specify)*								
Intraoperative (Neuro)								
Laparoscopic								
Pediatric	P	P					P	Note 1,2
Small organ(specify)**	P	P					P	Note 1,2
Neonatal Cephalic	P	P					P	Note 1,2
Adult Cephalic								
Trans-rectal								
Trans-vaginal								
Trans-urethral								
Trans-esoph.(non-Card.)								
Musculo-skeletal Conventional	P	P					P	Note 1,2
Musculo-skeletal Superficial	P	P					P	Note 1,2
Intravascular								
Cardiac Adult								
Cardiac Pediatric								
Intravascular (Cardiac)								
Trans-esoph.(Cardiac)								
Intra-Cardiac								
Peripheral Vascular	P	P					P	Note 1,2
Other (specify)***								

N=new indication; P=previously cleared by FDA; E=added under Appendix E

Additional comments: Combined modes: B+M.

*Intraoperative includes abdominal, thoracic, and vascular

**Small organ-breast, thyroid, testes

***Other use includes Urology.


Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: Biopsy Guidance

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)


 (Division Sign-Off)
 Division of Radiological Devices
 Office of In Vitro Diagnostic Device Evaluation and Safety
 510K K111435

Diagnostic Ultrasound Indications for Use Form

System: DP-50 Digital Ultrasonic Diagnostic Imaging System
 Transducer: 10L24EA
 Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation							
	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic								
Fetal								
Abdominal	N	N					N	Note 1,2
Intraoperative (specify)*								
Intraoperative (Neuro)								
Laparoscopic								
Pediatric	N	N					N	Note 1,2
Small organ(specify)**	N	N					N	Note 1,2
Neonatal Cephalic	N	N					N	Note 1,2
Adult Cephalic								
Trans-rectal								
Trans-vaginal								
Trans-urethral								
Trans-esoph.(non-Card.)								
Musculo-skeletal Conventional	N	N					N	Note 1,2
Musculo-skeletal Superficial	N	N					N	Note 1,2
Intravascular								
Cardiac Adult								
Cardiac Pediatric								
Intravascular (Cardiac)								
Trans-esoph.(Cardiac)								
Intra-Cardiac								
Peripheral Vascular	N	N					N	Note 1,2
Other (specify)***								

N=new indication; P=previously cleared by FDA; E=added under Appendix E

Additional comments: Combined modes: B+M.

*Intraoperative includes abdominal, thoracic, and vascular

**Small organ-breast, thyroid, testes

***Other use includes Urology.


Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note 2: Biopsy Guidance

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)


 (Division Sign-Off)
 Division of Radiological Devices
 Office of In Vitro Diagnostic Device Evaluation and Safety
 510K K111435

Diagnostic Ultrasound Indications for Use Form

System DP-50 Digital Ultrasonic Diagnostic Imaging System
 Transducer: 65EB10EA
 Intended Use: Diagnostic ultrasound imaging or fluid flow analysis of the human body as follows:

Clinical Application	Mode of Operation							
	B	M	PWD	CWD	Color Doppler	Amplitude Doppler	Combined (specify)	Other (specify)
Ophthalmic								
Fetal								
Abdominal								
Intraoperative (specify)*								
Intraoperative (Neuro)								
Laparoscopic								
Pediatric								
Small organ(specify)**								
Neonatal Cephalic								
Adult Cephalic								
Trans-rectal	N	N					N	Note 1,2
Trans-vaginal								
Trans-urethral								
Trans-esoph.(non-Card.)								
Musculo-skeletal Conventional								
Musculo-skeletal Superficial								
Intravascular								
Cardiac Adult								
Cardiac Pediatric								
Intravascular (Cardiac)								
Trans-esoph.(Cardiac)								
Intra-Cardiac								
Peripheral Vascular								
Other (specify)***								

N=new indication; P=previously cleared by FDA; E=added under Appendix E

Additional comments: Combined modes: B+M.

*Intraoperative includes abdominal, thoracic, and vascular

**Small organ-breast, thyroid, testes

***Other use includes Urology.


Note 1: Tissue Harmonic Imaging. The feature does not use contrast agents.

Note2: Biopsy Guidance

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation(ODE)

Prescription USE (Per 21 CFR 801.109)


 (Division Sign-Off)
 Division of Radiological Devices
 Office of In Vitro Diagnostic Device Evaluation and Safety

510K

K111435

0036